

## **Certification Record**

CUSTOMER	CLASS	FILE
SycoTec GmbH & Co KG	8721-85	241250
Wangener Strasse 78,	ELECTRICAL EQUIPMENT FOR LABORATORY USE-Certified to US Standards	
Leutkirch		
88299		
Germany	Refer to Class Description for program details	

- **Dental Maintenance Machine**; model STAT*MATIC*, type 31, cord connected, rated AC 100 230 V, 50/60 Hz,
  - 30 VA, Equipment Class I, Pollution Degree 2, Installation Category II
- Motors for laboratory applications, Models 4040 DC-S and 4041 DC-S, rated 45 V, 1000 Hz, 1.0 kW, 3-phase.

## Notes:

1. Devices are Certified as components for use only in other Certified equipment where the suitability of the combination is determined by CSA International.

/cc

Copyright © 2009 CSA International. All rights reserved.



## **Certification Record**

CUSTOMER	CLASS	FILE
SycoTec GmbH & Co KG	<u>8721-05</u>	241250
Wangener Strasse 78,	LABORATORY EQUIPMENT- Electrical	
Leutkirch		
88299		
Germany	Refer to Class Description for program details	

- **Dental Maintenance Machine;** model STAT*MATIC*, type 31, cord connected, rated AC 100 230 V, 50/60 Hz,
  - 30 VA, Equipment Class I, Pollution Degree 2, Installation Category II
- Motors for laboratory applications, Models 4040 DC-S and 4041 DC-S, rated 45 V, 1000 Hz, 1.0 kW, 3-phase.

## Notes:

 Devices are Certified as components for use only in other Certified equipment where the suitability of the combination is determined by CSA International.

/cc

Copyright © 2009 CSA International. All rights reserved.